

PH METER



DESCRIPTION:

To determine the concentration of Hydrogen ions in an aqueous solution.

HOT PLATE & MAGNETIC STIRRER



DESCRIPTION:

For homogenous stirring of solution

OVEN



DESCRIPTION:

Used for sterilization and sample preparation and drying.

CONDITIONING CHAMBER



DESCRIPTION:

For the conditioning of samples prior to testing, provides conditions from 20% to 98% RH \pm 2 and 0°C to 94°C \pm 0.5°C.

LAUNDROMETER



DESCRIPTION: To determine wash fastness of a given material

STANDARDS: AATCC61. ISO 105 etc..

VISCOMETER



DESCRIPTION: To determine the viscosity of a given solution

FOGGING TESTER



DESCRIPTION:

To determine the fogging characteristics evaluation of volatile constituents of decorating materials used in cars and aircrafts,

APPLICATION: Textiles, Leather, Foams and other materials.

STANDARDS: ISO 6452, DIN 75201, SAE J1756, QB/T 2728, BS EN 14288, PV 3920, PV 3015, ES-X83231, NES M0161, D45 1727, GM 9305P, TSM 0503G